



EU-TYPE EXAMINATION CERTIFICATE

No. ICR Polska/R6001119

Name and address of the certificate holder / manufacturer

Sentech Korea Corp.
21-6, Jimokro 75-gil, Gyeonggi-do, Paju-Si, 10880, Korea

Product name

Breathalyzer

Product type

ALP-1 / ALP-1 (SD), AMP100 Professionell (SD), ACE Pro Med (SD),
Alcodetector Pro1 (SD), Dingo E-200 B (SD), AlcoDigital EON (SD)

Conformity with the essential requirements was ensured by compliance with the following standards

<u>Essential requirements</u>	<u>Standard/specification</u>
Article 3.1.a Health	EN 62479:2010
Article 3.1.a Safety	EN 61010-1:2010
Article 3.1.b EMC	ETSI EN 301489-1 V2.1.1 (2017-02) ETSI EN 301489-17 V3.1.1 (2017-02) EN 61326-1:2013; EN 61326-2-3:2013
Article 3.2 Radio Spectrum	ETSI EN 300 328 V2.1.1 (2016-11)
Article 3.3 Special Radio Equipment Requirements	-

Test reports

TRSCES19-0016	related to article 3.1a
ICRT-TR-E190597-0A	related to article 3.1b
ICRT-TR-E190596-0A	related to article 3.1b
DRTCET1806-0272(1)	related to article 3.2
ICRT-TR-E190669-0A	related to article 3.2

Notified body no 2703

State that the conformity assessment has been carried out in accordance with the Program PC-P-07-05 and provided evidence of compliance with above scope of the 2014/53/EU.

It applies only to the sample submitted to ICR Polska Co. Ltd.

Date of certification: 27.06.2019

The mutual obligations and rights of the parties related to the subject matter of the certification are governed by the agreement no ICR Polska/2019-0096.

This certificate applies to products having the same characteristics (parameters) that have been evaluated and meet the requirements of the relevant standards.



AC 197



Director: Rafał Kalinowski

Warsaw, 27.06.2019

ICR Polska Sp. z o.o

ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrqa.com



Notified Body

CE
2703

